

Medical Policy



Lower Limb Prosthesis

▼ Description

A lower limb prosthesis or artificial limb is used following the amputation or removal of an injured or diseased lower limb. The prosthesis allows for normal ambulation.

▼ Policy

A lower limb prosthesis is **reasonable and necessary**, following the amputation or removal of an injured or diseased lower limb, to allow for normal ambulation.

▼ Policy Guidelines

Coverage Criteria:

1. Must be ordered by the Member's treating physician.
2. A lower limb prosthesis is covered when the Member:
 - a. Will reach or maintain a defined functional state within a reasonable period of time; and
 - b. Is motivated to ambulate
3. Functional levels:

A determination of the medical necessity for certain components/additions to the prosthesis is based on the Member's potential functional abilities. Potential functional ability is based on the reasonable expectations of the prosthetist, and treating physician, considering factors including, but not limited to:

 - a. The Member's past history (including prior prosthetic use if applicable); and
 - b. The Member's current condition including the status of the residual limb and the nature of other medical problems; and
 - c. The Member's desire to ambulate.
4. Clinical assessments of Member rehabilitation potential must be based on the following classification levels:

Level 0: Does not have the ability or potential to ambulate or transfer safely with or without assistance and a prosthesis does not enhance their quality of life or mobility.

Level 1: Has the ability or potential to use a prosthesis for transfers or ambulation on level surfaces at fixed cadence. Typical of the limited and unlimited household ambulator.

Level 2: Has the ability or potential for ambulation with the ability to traverse low level environmental barriers such as curbs, stairs or uneven surfaces. Typical of the limited community ambulator.

Level 3: Has the ability or potential for ambulation with variable cadence. Typical of the community ambulator who has the ability to traverse most environmental barriers and may have vocational, therapeutic, or exercise activity that demands prosthetic utilization beyond simple locomotion.

Level 4: Has the ability or potential for prosthetic ambulation that exceeds basic ambulation skills, exhibiting high impact, stress, or energy levels. Typical of the prosthetic demands of the child, active adult, or athlete.

The records must document the Member's current functional capabilities and his/her expected functional potential, including an explanation for the difference, if that is the case. It is recognized, within the functional classification hierarchy, that bilateral amputees often cannot be strictly bound by functional level classifications.

5. General:

When an initial below knee prosthesis (L5500) or a preparatory below knee prosthesis (L5510-L5530, L5540) is provided, prosthetic substitutions and/or additions of procedures and components are covered in accordance with the functional level assessment except for codes L5629, L5638, L5639, L5646, L5647, L5704, L5785, L5962, and L5980 which will be denied as not medically necessary. When a below knee preparatory prefabricated prosthesis (L5535) is provided, prosthetic substitutions and/or additions of procedures are covered in accordance with the functional level assessment except for codes L5620, L5629, L5645, L5646, L5670, L5676, L5704, and L5962 which will be denied as not medically necessary.

When an above knee initial prosthesis (L5505) or an above knee preparatory (L5560-L5580, L5590-L5600) prosthesis is provided, prosthetic substitution and/or additions of procedures and components are covered in accordance with the functional level assessment except for codes L5610, L5631, L5640, L5642, L5644, L5648, L5705, L5706, L5964, L5980, and L5710-L5780, L5790-L5795 which will be denied as not medically necessary. When an above knee preparatory prefabricated prosthesis (L5585) is provided, prosthetic substitution and/or additions of procedures and components are covered in accordance with the functional level assessment except for codes L5624, L5631, L5648,

L5651, L5652, L5705, L5706, L5964, and L5966 which will be denied as not medically necessary.

Lower limb prosthetic covers (L5704-L5707) are complete products and afford shape, protection and waterproofing for normal daily usage of the prosthesis. They offer sufficient protection and weatherproofing for beneficiaries who require lower limb prosthetics.

Protective outer surface covering systems (L5962, L5964, and L5966) are specialized covers intended to be worn over an existing prosthesis. They are used by a beneficiary who has special needs for protection against unusually harsh environmental situations where it is necessary to protect the lower limb prosthesis beyond the level of protection that is afforded by L5704-L5707. They are not for cosmetic or convenience reasons, or for everyday usage in a typical environment. Protective outer surface coverings are different from the covering that is already reimbursed as part of L5704 – L5707.

In the following sections, the determination of coverage for selected prostheses and components with respect to potential functional levels represents the usual case. Exceptions will be considered in an individual case if additional documentation is included which justifies the medical necessity. Prostheses will be denied as not medically necessary if the Member's potential functional level is 0.

FEET:

A determination of the type of foot for the prosthesis will be made by the treating physician and/or the prosthetist based upon the functional needs of the Member. Basic lower extremity prostheses include a SACH foot. Other prosthetic feet are considered for coverage based upon functional classification.

- An external keel SACH foot (L5970) or single axis ankle/foot (L5974) is covered for Members whose functional level is 1 or above.

- A flexible-keel foot (L5972) or multi-axial ankle/foot (L5978) is covered for Members whose functional level is 2 or above.

- A microprocessor controlled ankle foot system (L5973), energy storing foot (L5976), dynamic response foot with multi-axial ankle (L5979), flex foot system (L5980), flex-walk system or equal (L5981), or shank foot system with vertical loading pylon (L5987) is covered for members whose functional level is 3 or above.

The microprocessor foot or ankle system addition with power assist which includes any type motor (L5969), is not considered because there is

insufficient information to demonstrate that the item meets the standard to be considered reasonable and necessary.

Coverage is extended only if there is sufficient clinical documentation of functional need for the technologic or design feature of a given type of foot. This information must be retained in the physician's or prosthetist's files.

A user adjustable heel height feature (L5990) will be denied as not reasonable and necessary.

KNEES:

A determination of the type of knee for the prosthesis will be made by the treating physician and/or the prosthetist based upon the functional needs of the Member. Basic lower extremity prostheses include a single axis, constant friction knee. Other prosthetic knees are considered for coverage based upon functional classification.

- A high activity knee control frame (L5930) is covered for Members whose functional level is 4.

- A fluid, pneumatic, or electronic knee (L5610, L5613, L5614, L5722 - L5780, L5814, L5822-L5840, L5848, L5856, L5857, L5858) is covered for Members whose functional level is 3 or above.

L5859 (ADDITION TO LOWER EXTREMITY PROSTHESIS, ENDOSKELETAL KNEE-SHIN SYSTEM, POWERED AND PROGRAMMABLE FLEXION/EXTENSION ASSIST CONTROL, INCLUDES ANY TYPE MOTOR(S)) is only covered when the beneficiary meets all of the criteria below:

1. Has a microprocessor (swing and stance phase type (L5856)) controlled (electronic) knee
2. K3 functional level only
3. Has a documented comorbidity of the spine and/or sound limb affecting hip extension and/or quadriceps function that impairs K-3 level function with the use of a microprocessor-controlled knee alone
4. Is able to make use of a product that requires daily charging
5. Is able to understand and respond to error alerts and alarms indicating problems with the function of the unit

If these coverage criteria for the knee component are not met, L5859 will be denied as not reasonable and necessary.

Other knee systems (L5611, L5616, L5710-L5718, L5810-L5812, L5816,

L5818) are covered for Members whose functional level is 1 or above.

Coverage is extended only if there is sufficient clinical documentation of functional need for the technologic or design feature of a given type of knee. This information must be retained in the physician's or prosthetist's files.

ANKLES:

An axial rotation unit (L5982-L5986) is covered for Members whose functional level is 2 or above.

SOCKETS:

More than 2 test (diagnostic) sockets (L5618-L5628) for an individual prosthesis are not medically necessary unless there is documentation in the medical record which justifies the need. Exception: A test socket is not medically necessary for an immediate prosthesis (L5400-L5460).

No more than two of the same socket inserts (L5654-L5665, L5673, L5679, L5681, L5683) are allowed per individual prosthesis at the same time.

Socket replacements are considered medically necessary if there is adequate documentation of functional and/or physiological need. It is recognized that there are situations where the explanation includes but is not limited to: changes in the residual limb; functional need changes; or irreparable damage or wear/tear due to excessive Member weight or prosthetic demands of very active amputees.

Powered base items are those that contain the power source (battery). At the time that a base item is billed, all necessary batteries and/or battery chargers are considered as included in the payment for the powered base item. There is no separate payment for batteries (L7360, L7364, L7367, and L8505) and/or battery chargers (L7362, L7366, L7368) billed concurrently with a powered base item.

Payments for items listed in Column II are included in the payment for each Column I code. Claims for Column II items billed with the provision of a Column I item will be denied as unbundling.

Column I

Base codes with battery, charger and/or power included
L5781
L5782
L5856
L5857

Column II

Batteries
L7360
L7364
L7367
L8505

L5858
L5859
L5973

Chargers
L7362
L7366
L7368

Limitations:

1. The following items are included in the reimbursement for a prosthesis and therefore, are not separately billable:
 - a. Repairs due to normal wear or tear within 90 days of delivery.
 - b. Adjustments of the prosthesis or the prosthetic component made when fitting the prosthesis or component and for 90 days from the date of delivery when the adjustments are not necessitated by changes in the residual limb or the Member's functional abilities.
 - c. Evaluation of the residual limb and gait.
 - d. Fitting of the prosthesis.
 - e. Costs of base component parts and labor contained in HCPCS base codes.
2. The replacement of artificial limbs are covered if any of the following criteria is met:
 - a. A change in the physical condition of the Member.
 - b. An irreparable change in the condition of the device, or in a part of the device.
 - c. The condition of the device, or part of the device, requires repairs and the cost of such repairs would be more than sixty (60) percent of the cost of a replacement device, or, of the part being replaced.
3. Custom fabricated socket inserts for congenital or atypical amputee (L5681, L5683) will only be covered when the initial prosthesis is requested.
4. No more than 2 test (diagnostic) sockets are allowed per individual prosthesis.
5. No more than two of the same socket inserts (L5654-L5658, L5661, L5665, L5673, L5679, L5681, L5683) are allowed per individual prosthesis at the same time.
6. Code L7700 (GASKET OR SEAL, FOR USE WITH PROSTHETIC SOCKET INSERT, ANY TYPE, EACH) describes a stand-alone (i.e., not integrated into or a part of a prosthetic socket insert) sealing ring that is added to a socket insert to assist in providing or maintaining negative pressure for socket suspension. The ring creates a seal against the outer surface of the insert and against the inner wall of the socket. L7700 is not intended for use with mechanical socket suspensions such as a pin-lock system. It may be made of any suitable material. L7700 may be used with upper or lower extremity sockets. Unit of service (UOS) is 1 (one) item. This code is not to be used to bill for gaskets, seals, or other sealing materials that are included as part of an insert. Integrated seals are

included in the code for the insert. Separate billing of integrated gaskets or seals as L7700 is unbundling.

7. Accessories are limited to the quantities listed below:

HCPCS Code	Description	Quantity
L8400	Prosthetic sheath, below knee, each	Up to 12 in a 12 month period
L8410	Prosthetic sheath, above knee, each	Up to 12 in a 12 month period
L8417	Prosthetic sheath/sock, including a gel cushion layer, below or above knee, each	Up to 12 in a 12 month period
L8420	Prosthetic sock, multiple ply, below knee, each	Up to 12 in a 12 month period
L8430	Prosthetic sock, multiple ply, above knee, each	Up to 12 in a 12 month period
L8440	Prosthetic shrinker, below knee, each	Up to 4 in a 12 month period
L8460	Prosthetic shrinker, above knee, each	Up to 4 in a 12 month period
L8470	Stump sock, single ply, below knee, each	Up to 12 in a 12 month period
L8480	Stump sock, single ply, above knee, each	Up to 12 in a 12 month period

Exclusions:

1. If a prosthesis is denied as not medically necessary, related additions will also be denied as not medically necessary.
2. Immediate post-surgical or early fitting procedures (IPOP), procedure codes (L5400, L5410, L5420, L5430, L5450 and L5460) are normally provided to a Member while in a hospital stay and are not covered at home.

▼ HCPCS Level II Codes and Description

L5000 PARTIAL FOOT, SHOE INSERT WITH LONGITUDINAL ARCH, TOE FILLER

L5010 PARTIAL FOOT, MOLDED SOCKET, ANKLE HEIGHT, WITH TOE FILLER

L5020 PARTIAL FOOT, MOLDED SOCKET, TIBIAL TUBERCLE HEIGHT, WITH TOE FILLER

L5050 ANKLE, SYMES, MOLDED SOCKET, SACH FOOT

L5060 ANKLE, SYMES, METAL FRAME, MOLDED LEATHER SOCKET, ARTICULATED ANKLE/FOOT

L5100 BELOW KNEE, MOLDED SOCKET, SHIN, SACH FOOT

L5105 BELOW KNEE, PLASTIC SOCKET, JOINTS AND THIGH LACER, SACH FOOT

L5150 KNEE DISARTICULATION (OR THROUGH KNEE), MOLDED SOCKET, EXTERNAL KNEE JOINTS, SHIN, SACH FOOT

L5160 KNEE DISARTICULATION (OR THROUGH KNEE), MOLDED SOCKET, BENT KNEE CONFIGURATION, EXTERNAL KNEE JOINTS, SHIN, SACH FOOT

L5200 ABOVE KNEE, MOLDED SOCKET, SINGLE AXIS CONSTANT FRICTION KNEE, SHIN, SACH FOOT

L5210 ABOVE KNEE, SHORT PROSTHESIS, NO KNEE JOINT ('STUBBIES'), WITH FOOT BLOCKS, NO ANKLE JOINTS, EACH

L5220 ABOVE KNEE, SHORT PROSTHESIS, NO KNEE JOINT ('STUBBIES'), WITH ARTICULATED ANKLE/FOOT, DYNAMICALLY ALIGNED, EACH

L5230 ABOVE KNEE, FOR PROXIMAL FEMORAL FOCAL DEFICIENCY, CONSTANT FRICTION KNEE, SHIN, SACH FOOT

L5250 HIP DISARTICULATION, CANADIAN TYPE; MOLDED SOCKET, HIP JOINT, SINGLE AXIS CONSTANT FRICTION KNEE, SHIN, SACH FOOT

L5270 HIP DISARTICULATION, TILT TABLE TYPE; MOLDED SOCKET, LOCKING HIP JOINT, SINGLE AXIS CONSTANT FRICTION KNEE, SHIN, SACH FOOT

L5280 HEMIPELVECTOMY, CANADIAN TYPE; MOLDED SOCKET, HIP JOINT, SINGLE AXIS CONSTANT FRICTION KNEE, SHIN, SACH FOOT

L5301 BELOW KNEE, MOLDED SOCKET, SHIN, SACH FOOT, ENDOSKELETAL SYSTEM

L5312 KNEE DISARTICULATION (OR THROUGH KNEE), MOLDED SOCKET, SINGLE AXIS KNEE, PYLON, SACH FOOT, ENDOSKELETAL SYSTEM.

L5321 ABOVE KNEE, MOLDED SOCKET, OPEN END, SACH FOOT,

ENDOSKELETAL SYSTEM, SINGLE AXIS KNEE

L5331 HIP DISARTICULATION, CANADIAN TYPE, MOLDED SOCKET, ENDOSKELETAL SYSTEM, HIP JOINT, SINGLE AXIS KNEE, SACH FOOT

L5341 HEMIPELVECTOMY, CANADIAN TYPE, MOLDED SOCKET, ENDOSKELETAL SYSTEM, HIP JOINT, SINGLE AXIS KNEE, SACH FOOT

L5400 IMMEDIATE POST SURGICAL OR EARLY FITTING, APPLICATION OF INITIAL RIGID DRESSING, INCLUDING FITTING, ALIGNMENT, SUSPENSION, AND ONE CAST CHANGE, BELOW KNEE

L5410 IMMEDIATE POST SURGICAL OR EARLY FITTING, APPLICATION OF INITIAL RIGID DRESSING, INCLUDING FITTING, ALIGNMENT AND SUSPENSION, BELOW KNEE, EACH ADDITIONAL CAST CHANGE AND REALIGNMENT

L5420 IMMEDIATE POST SURGICAL OR EARLY FITTING, APPLICATION OF INITIAL RIGID DRESSING, INCLUDING FITTING, ALIGNMENT AND SUSPENSION AND ONE CAST CHANGE 'AK' OR KNEE DISARTICULATION

L5430 IMMEDIATE POST SURGICAL OR EARLY FITTING, APPLICATION OF INITIAL RIGID DRESSING, INCL. FITTING, ALIGNMENT AND SUSPENSION, 'AK' OR KNEE DISARTICULATION, EACH ADDITIONAL CAST CHANGE AND REALIGNMENT

L5450 IMMEDIATE POST SURGICAL OR EARLY FITTING, APPLICATION OF NON-WEIGHT BEARING RIGID DRESSING, BELOW KNEE

L5460 IMMEDIATE POST SURGICAL OR EARLY FITTING, APPLICATION OF NON-WEIGHT BEARING RIGID DRESSING, ABOVE KNEE

L5500 INITIAL, BELOW KNEE 'PTB' TYPE SOCKET, NON-ALIGNABLE SYSTEM, PYLON, NO COVER, SACH FOOT, PLASTER SOCKET, DIRECT FORMED

L5505 INITIAL, ABOVE KNEE - KNEE DISARTICULATION, ISCHIAL LEVEL SOCKET, NON-ALIGNABLE SYSTEM, PYLON, NO COVER, SACH FOOT, PLASTER SOCKET, DIRECT FORMED

L5510 PREPARATORY, BELOW KNEE 'PTB' TYPE SOCKET, NON-ALIGNABLE SYSTEM, PYLON, NO COVER, SACH FOOT, PLASTER SOCKET, MOLDED TO MODEL

L5520 PREPARATORY, BELOW KNEE 'PTB' TYPE SOCKET, NON-ALIGNABLE SYSTEM, PYLON, NO COVER, SACH FOOT, THERMOPLASTIC OR EQUAL, DIRECT FORMED

L5530 PREPARATORY, BELOW KNEE 'PTB' TYPE SOCKET, NON-

ALIGNABLE SYSTEM, PYLON, NO COVER, SACH FOOT,
THERMOPLASTIC OR EQUAL, MOLDED TO MODEL

L5535 PREPARATORY, BELOW KNEE 'PTB' TYPE SOCKET, NON-
ALIGNABLE SYSTEM, NO COVER, SACH FOOT, PREFABRICATED,
ADJUSTABLE OPEN END SOCKET

L5540 PREPARATORY, BELOW KNEE 'PTB' TYPE SOCKET, NON-
ALIGNABLE SYSTEM, PYLON, NO COVER, SACH FOOT, LAMINATED
SOCKET, MOLDED TO MODEL

L5560 PREPARATORY, ABOVE KNEE- KNEE DISARTICULATION, ISCHIAL
LEVEL SOCKET, NON-ALIGNABLE SYSTEM, PYLON, NO COVER,
SACH FOOT, PLASTER SOCKET, MOLDED TO MODEL

L5570 PREPARATORY, ABOVE KNEE - KNEE DISARTICULATION, ISCHIAL
LEVEL SOCKET, NON-ALIGNABLE SYSTEM, PYLON, NO COVER,
SACH FOOT, THERMOPLASTIC OR EQUAL, DIRECT FORMED

L5580 PREPARATORY, ABOVE KNEE - KNEE DISARTICULATION ISCHIAL
LEVEL SOCKET, NON-ALIGNABLE SYSTEM, PYLON, NO COVER,
SACH FOOT, THERMOPLASTIC OR EQUAL, MOLDED TO MODEL

L5585 PREPARATORY, ABOVE KNEE - KNEE DISARTICULATION, ISCHIAL
LEVEL SOCKET, NON-ALIGNABLE SYSTEM, PYLON, NO COVER,
SACH FOOT, PREFABRICATED ADJUSTABLE OPEN END SOCKET

L5590 PREPARATORY, ABOVE KNEE - KNEE DISARTICULATION ISCHIAL
LEVEL SOCKET, NON-ALIGNABLE SYSTEM, PYLON NO COVER,
SACH FOOT, LAMINATED SOCKET, MOLDED TO MODEL

L5595 PREPARATORY, HIP DISARTICULATION-HEMIPELVECTOMY,
PYLON, NO COVER, SACH FOOT, THERMOPLASTIC OR EQUAL,
MOLDED TO PATIENT MODEL

L5600 PREPARATORY, HIP DISARTICULATION-HEMIPELVECTOMY,
PYLON, NO COVER, SACH FOOT, LAMINATED SOCKET, MOLDED
TO PATIENT MODEL

L5610 ADDITION TO LOWER EXTREMITY, ENDOSKELETAL SYSTEM,
ABOVE KNEE, HYDRACADENCE SYSTEM

L5611 ADDITION TO LOWER EXTREMITY, ENDOSKELETAL SYSTEM,
ABOVE KNEE - KNEE DISARTICULATION, 4 BAR LINKAGE, WITH
FRICTION SWING PHASE CONTROL

L5613 ADDITION TO LOWER EXTREMITY, ENDOSKELETAL SYSTEM,
ABOVE KNEE-KNEE DISARTICULATION, 4 BAR LINKAGE, WITH
HYDRAULIC SWING PHASE CONTROL

L5614 ADDITION TO LOWER EXTREMITY, EXOSKELETAL SYSTEM, ABOVE

KNEE-KNEE DISARTICULATION, 4 BAR LINKAGE, WITH PNEUMATIC SWING PHASE CONTROL

- L5616 ADDITION TO LOWER EXTREMITY, ENDOSKELETAL SYSTEM, ABOVE KNEE, UNIVERSAL MULTIPLEX SYSTEM, FRICTION SWING PHASE CONTROL
- L5617 ADDITION TO LOWER EXTREMITY, QUICK CHANGE SELF-ALIGNING UNIT, ABOVE KNEE OR BELOW KNEE, EACH
- L5618 ADDITION TO LOWER EXTREMITY, TEST SOCKET, SYMES
- L5620 ADDITION TO LOWER EXTREMITY, TEST SOCKET, BELOW KNEE
- L5622 ADDITION TO LOWER EXTREMITY, TEST SOCKET, KNEE DISARTICULATION
- L5624 ADDITION TO LOWER EXTREMITY, TEST SOCKET, ABOVE KNEE
- L5626 ADDITION TO LOWER EXTREMITY, TEST SOCKET, HIP DISARTICULATION
- L5628 ADDITION TO LOWER EXTREMITY, TEST SOCKET, HEMIPELVECTOMY
- L5629 ADDITION TO LOWER EXTREMITY, BELOW KNEE, ACRYLIC SOCKET
- L5630 ADDITION TO LOWER EXTREMITY, SYMES TYPE, EXPANDABLE WALL SOCKET
- L5631 ADDITION TO LOWER EXTREMITY, ABOVE KNEE OR KNEE DISARTICULATION, ACRYLIC SOCKET
- L5632 ADDITION TO LOWER EXTREMITY, SYMES TYPE, 'PTB' BRIM DESIGN SOCKET
- L5634 ADDITION TO LOWER EXTREMITY, SYMES TYPE, POSTERIOR OPENING (CANADIAN) SOCKET
- L5636 ADDITION TO LOWER EXTREMITY, SYMES TYPE, MEDIAL OPENING SOCKET
- L5637 ADDITION TO LOWER EXTREMITY, BELOW KNEE, TOTAL CONTACT
- L5638 ADDITION TO LOWER EXTREMITY, BELOW KNEE, LEATHER SOCKET
- L5639 ADDITION TO LOWER EXTREMITY, BELOW KNEE, WOOD SOCKET
- L5640 ADDITION TO LOWER EXTREMITY, KNEE DISARTICULATION, LEATHER SOCKET
- L5642 ADDITION TO LOWER EXTREMITY, ABOVE KNEE, LEATHER SOCKET

- L5643 ADDITION TO LOWER EXTREMITY, HIP DISARTICULATION, FLEXIBLE INNER SOCKET, EXTERNAL FRAME
- L5644 ADDITION TO LOWER EXTREMITY, ABOVE KNEE, WOOD SOCKET
- L5645 ADDITION TO LOWER EXTREMITY, BELOW KNEE, FLEXIBLE INNER SOCKET, EXTERNAL FRAME
- L5646 ADDITION TO LOWER EXTREMITY, BELOW KNEE, AIR, FLUID, GEL OR EQUAL, CUSHION SOCKET
- L5647 ADDITION TO LOWER EXTREMITY, BELOW KNEE SUCTION SOCKET
- L5648 ADDITION TO LOWER EXTREMITY, ABOVE KNEE, AIR, FLUID, GEL OR EQUAL, CUSHION SOCKET
- L5649 ADDITION TO LOWER EXTREMITY, ISCHIAL CONTAINMENT/NARROW M-L SOCKET
- L5650 ADDITIONS TO LOWER EXTREMITY, TOTAL CONTACT, ABOVE KNEE OR KNEE DISARTICULATION SOCKET
- L5651 ADDITION TO LOWER EXTREMITY, ABOVE KNEE, FLEXIBLE INNER SOCKET, EXTERNAL FRAME
- L5652 ADDITION TO LOWER EXTREMITY, SUCTION SUSPENSION, ABOVE KNEE OR KNEE DISARTICULATION SOCKET
- L5653 ADDITION TO LOWER EXTREMITY, KNEE DISARTICULATION, EXPANDABLE WALL SOCKET
- L5654 ADDITION TO LOWER EXTREMITY, SOCKET INSERT, SYMES, (KEMBLO, PELITE, ALIPLAST, PLASTAZOTE OR EQUAL)
- L5655 ADDITION TO LOWER EXTREMITY, SOCKET INSERT, BELOW KNEE (KEMBLO, PELITE, ALIPLAST, PLASTAZOTE OR EQUAL)
- L5656 ADDITION TO LOWER EXTREMITY, SOCKET INSERT, KNEE DISARTICULATION (KEMBLO, PELITE, ALIPLAST, PLASTAZOTE OR EQUAL)
- L5658 ADDITION TO LOWER EXTREMITY, SOCKET INSERT, ABOVE KNEE (KEMBLO, PELITE, ALIPLAST, PLASTAZOTE OR EQUAL)
- L5661 ADDITION TO LOWER EXTREMITY, SOCKET INSERT, MULTI-DUROMETER SYMES
- L5665 ADDITION TO LOWER EXTREMITY, SOCKET INSERT, MULTI-DUROMETER, BELOW KNEE
- L5666 ADDITION TO LOWER EXTREMITY, BELOW KNEE, CUFF SUSPENSION

- L5668 ADDITION TO LOWER EXTREMITY, BELOW KNEE, MOLDED DISTAL CUSHION
- L5670 ADDITION TO LOWER EXTREMITY, BELOW KNEE, MOLDED SUPRACONDYLAR SUSPENSION ('PTS' OR SIMILAR)
- L5671 ADDITION TO LOWER EXTREMITY, BELOW KNEE / ABOVE KNEE SUSPENSION LOCKING MECHANISM (SHUTTLE, LANYARD OR EQUAL), EXCLUDES SOCKET INSERT
- L5672 ADDITION TO LOWER EXTREMITY, BELOW KNEE, REMOVABLE MEDIAL BRIM SUSPENSION
- L5673 ADDITION TO LOWER EXTREMITY, BELOW KNEE/ABOVE KNEE, CUSTOM FABRICATED FROM EXISTING MOLD OR PREFABRICATED, SOCKET INSERT, SILICONE GEL, ELASTOMERIC OR EQUAL, FOR USE WITH LOCKING MECHANISM
- L5676 ADDITIONS TO LOWER EXTREMITY, BELOW KNEE, KNEE JOINTS, SINGLE AXIS, PAIR
- L5677 ADDITIONS TO LOWER EXTREMITY, BELOW KNEE, KNEE JOINTS, POLYCENTRIC, PAIR
- L5678 ADDITIONS TO LOWER EXTREMITY, BELOW KNEE, JOINT COVERS, PAIR
- L5679 ADDITION TO LOWER EXTREMITY, BELOW KNEE/ABOVE KNEE, CUSTOM FABRICATED FROM EXISTING MOLD OR PREFABRICATED, SOCKET INSERT, SILICONE GEL, ELASTOMERIC OR EQUAL, NOT FOR USE WITH LOCKING MECHANISM
- L5680 ADDITION TO LOWER EXTREMITY, BELOW KNEE, THIGH LACER, NONMOLDED
- L5681 ADDITION TO LOWER EXTREMITY, BELOW KNEE/ABOVE KNEE, CUSTOM FABRICATED SOCKET INSERT FOR CONGENITAL OR ATYPICAL TRAUMATIC AMPUTEE, SILICONE GEL, ELASTOMERIC OR EQUAL, FOR USE WITH OR WITHOUT LOCKING MECHANISM, INITIAL ONLY (FOR OTHER THAN INITIAL, USE CODE L5673 OR L5679)
- L5682 ADDITION TO LOWER EXTREMITY, BELOW KNEE, THIGH LACER, GLUTEAL/ISCHIAL, MOLDED
- L5683 ADDITION TO LOWER EXTREMITY, BELOW KNEE/ABOVE KNEE, CUSTOM FABRICATED SOCKET INSERT FOR OTHER THAN CONGENITAL OR ATYPICAL TRAUMATIC AMPUTEE, SILICONE GEL, ELASTOMERIC OR EQUAL, FOR USE WITH OR WITHOUT LOCKING MECHANISM, INITIAL ONLY (FOR OTHER THAN INITIAL, USE CODE L5673 OR L5679)

- L5684 ADDITION TO LOWER EXTREMITY, BELOW KNEE, FORK STRAP
- L5685 ADDITION TO LOWER EXTREMITY PROSTHESIS, BELOW KNEE, SUSPENSION/SEALING SLEEVE, WITH OR WITHOUT VALVE, ANY MATERIAL, EACH
- L5686 ADDITION TO LOWER EXTREMITY, BELOW KNEE, BACK CHECK (EXTENSION CONTROL)
- L5688 ADDITION TO LOWER EXTREMITY, BELOW KNEE, WAIST BELT, WEBBING
- L5690 ADDITION TO LOWER EXTREMITY, BELOW KNEE, WAIST BELT, PADDED AND LINED
- L5692 ADDITION TO LOWER EXTREMITY, ABOVE KNEE, PELVIC CONTROL BELT, LIGHT
- L5694 ADDITION TO LOWER EXTREMITY, ABOVE KNEE, PELVIC CONTROL BELT, PADDED AND LINED
- L5695 ADDITION TO LOWER EXTREMITY, ABOVE KNEE, PELVIC CONTROL, SLEEVE SUSPENSION, NEOPRENE OR EQUAL, EACH
- L5696 ADDITION TO LOWER EXTREMITY, ABOVE KNEE OR KNEE DISARTICULATION, PELVIC JOINT
- L5697 ADDITION TO LOWER EXTREMITY, ABOVE KNEE OR KNEE DISARTICULATION, PELVIC BAND
- L5698 ADDITION TO LOWER EXTREMITY, ABOVE KNEE OR KNEE DISARTICULATION, SILESIA BANDAGE
- L5699 ALL LOWER EXTREMITY PROSTHESES, SHOULDER HARNESS
- L5700 REPLACEMENT, SOCKET, BELOW KNEE, MOLDED TO PATIENT MODEL
- L5701 REPLACEMENT, SOCKET, ABOVE KNEE/KNEE DISARTICULATION, INCLUDING ATTACHMENT PLATE, MOLDED TO PATIENT MODEL
- L5702 REPLACEMENT, SOCKET, HIP DISARTICULATION, INCLUDING HIP JOINT, MOLDED TO PATIENT MODEL
- L5703 ANKLE, SYMES, MOLDED TO PATIENT MODEL, SOCKET WITHOUT SOLID ANKLE CUSHION HEEL (SACH) FOOT, REPLACEMENT ONLY
- L5704 CUSTOM SHAPED PROTECTIVE COVER, BELOW KNEE
- L5705 CUSTOM SHAPED PROTECTIVE COVER, ABOVE KNEE
- L5706 CUSTOM SHAPED PROTECTIVE COVER, KNEE DISARTICULATION
- L5707 CUSTOM SHAPED PROTECTIVE COVER, HIP DISARTICULATION

- L5710 ADDITION, EXOSKELETAL KNEE-SHIN SYSTEM, SINGLE AXIS, MANUAL LOCK
- L5711 ADDITIONS EXOSKELETAL KNEE-SHIN SYSTEM, SINGLE AXIS, MANUAL LOCK, ULTRA-LIGHT MATERIAL
- L5712 ADDITION, EXOSKELETAL KNEE-SHIN SYSTEM, SINGLE AXIS, FRICTION SWING AND STANCE PHASE CONTROL (SAFETY KNEE)
- L5714 ADDITION, EXOSKELETAL KNEE-SHIN SYSTEM, SINGLE AXIS, VARIABLE FRICTION SWING PHASE CONTROL
- L5716 ADDITION, EXOSKELETAL KNEE-SHIN SYSTEM, POLYCENTRIC, MECHANICAL STANCE PHASE LOCK
- L5718 ADDITION, EXOSKELETAL KNEE-SHIN SYSTEM, POLYCENTRIC, FRICTION SWING AND STANCE PHASE CONTROL
- L5722 ADDITION, EXOSKELETAL KNEE-SHIN SYSTEM, SINGLE AXIS, PNEUMATIC SWING, FRICTION STANCE PHASE CONTROL
- L5724 ADDITION, EXOSKELETAL KNEE-SHIN SYSTEM, SINGLE AXIS, FLUID SWING PHASE CONTROL
- L5726 ADDITION, EXOSKELETAL KNEE-SHIN SYSTEM, SINGLE AXIS, EXTERNAL JOINTS FLUID SWING PHASE CONTROL
- L5728 ADDITION, EXOSKELETAL KNEE-SHIN SYSTEM, SINGLE AXIS, FLUID SWING AND STANCE PHASE CONTROL
- L5780 ADDITION, EXOSKELETAL KNEE-SHIN SYSTEM, SINGLE AXIS, PNEUMATIC/HYDRA PNEUMATIC SWING PHASE CONTROL
- L5781 ADDITION TO LOWER LIMB PROSTHESIS, VACUUM PUMP, RESIDUAL LIMB VOLUME MANAGEMENT AND MOISTURE EVACUATION SYSTEM
- L5782 ADDITION TO LOWER LIMB PROSTHESIS, VACUUM PUMP, RESIDUAL LIMB VOLUME MANAGEMENT AND MOISTURE EVACUATION SYSTEM, HEAVY DUTY
- L5785 ADDITION, EXOSKELETAL SYSTEM, BELOW KNEE, ULTRA-LIGHT MATERIAL (TITANIUM, CARBON FIBER OR EQUAL)
- L5790 ADDITION, EXOSKELETAL SYSTEM, ABOVE KNEE, ULTRA-LIGHT MATERIAL (TITANIUM, CARBON FIBER OR EQUAL)
- L5795 ADDITION, EXOSKELETAL SYSTEM, HIP DISARTICULATION, ULTRA-LIGHT MATERIAL (TITANIUM, CARBON FIBER OR EQUAL)
- L5810 ADDITION, ENDOSKELETAL KNEE-SHIN SYSTEM, SINGLE AXIS, MANUAL LOCK

- L5811 ADDITION, ENDOSKELETAL KNEE-SHIN SYSTEM, SINGLE AXIS, MANUAL LOCK, ULTRA-LIGHT MATERIAL
- L5812 ADDITION, ENDOSKELETAL KNEE-SHIN SYSTEM, SINGLE AXIS, FRICTION SWING AND STANCE PHASE CONTROL (SAFETY KNEE)
- L5814 ADDITION, ENDOSKELETAL KNEE-SHIN SYSTEM, POLYCENTRIC, HYDRAULIC SWING PHASE CONTROL, MECHANICAL STANCE PHASE LOCK
- L5816 ADDITION, ENDOSKELETAL KNEE-SHIN SYSTEM, POLYCENTRIC, MECHANICAL STANCE PHASE LOCK
- L5818 ADDITION, ENDOSKELETAL KNEE-SHIN SYSTEM, POLYCENTRIC, FRICTION SWING, AND STANCE PHASE CONTROL
- L5822 ADDITION, ENDOSKELETAL KNEE-SHIN SYSTEM, SINGLE AXIS, PNEUMATIC SWING, FRICTION STANCE PHASE CONTROL
- L5824 ADDITION, ENDOSKELETAL KNEE-SHIN SYSTEM, SINGLE AXIS, FLUID SWING PHASE CONTROL
- L5826 ADDITION, ENDOSKELETAL KNEE-SHIN SYSTEM, SINGLE AXIS, HYDRAULIC SWING PHASE CONTROL, WITH MINIATURE HIGH ACTIVITY FRAME
- L5828 ADDITION, ENDOSKELETAL KNEE-SHIN SYSTEM, SINGLE AXIS, FLUID SWING AND STANCE PHASE CONTROL
- L5830 ADDITION, ENDOSKELETAL KNEE-SHIN SYSTEM, SINGLE AXIS, PNEUMATIC/ SWING PHASE CONTROL
- L5840 ADDITION, ENDOSKELETAL KNEE/SKIN SYSTEM, 4-BAR LINKAGE OR MULTIAXIAL, PNEUMATIC SWING PHASE CONTROL
- L5845 ADDITION, ENDOSKELETAL, KNEE-SHIN SYSTEM, STANCE FLEXION FEATURE, ADJUSTABLE
- L5848 ADDITION TO ENDOSKELETAL KNEE-SHIN SYSTEM, FLUID STANCE EXTENSION, DAMPENING FEATURE, WITH OR WITHOUT ADJUSTABILITY
- L5850 ADDITION, ENDOSKELETAL SYSTEM, ABOVE KNEE OR HIP DISARTICULATION, KNEE EXTENSION ASSIST
- L5855 ADDITION, ENDOSKELETAL SYSTEM, HIP DISARTICULATION, MECHANICAL HIP EXTENSION ASSIST
- L5856 ADDITION TO LOWER EXTREMITY PROSTHESIS, ENDOSKELETAL KNEE-SHIN SYSTEM, MICROPROCESSOR CONTROL FEATURE, SWING AND STANCE PHASE, INCLUDES ELECTRONIC SENSOR(S), ANY TYPE

- L5857 ADDITION TO LOWER EXTREMITY PROSTHESIS, ENDOSKELETAL KNEE-SHIN SYSTEM, MICROPROCESSOR CONTROL FEATURE, SWING PHASE ONLY, INCLUDES ELECTRONIC SENSOR(S), ANY TYPE
- L5858 ADDITION TO LOWER EXTREMITY PROSTHESIS, ENDOSKELETAL KNEE SHIN SYSTEM, MICROPROCESSOR CONTROL FEATURE, STANCE PHASE ONLY, INCLUDES ELECTRONIC SENSOR(S), ANY TYPE
- L5859 ADDITION TO LOWER EXTREMITY PROSTHESIS, ENDOSKELETAL KNEE-SHIN SYSTEM, POWERED AND PROGRAMMABLE FLEXION/EXTENSION ASSIST CONTROL, INCLUDES ANY TYPE MOTOR(S)
- L5910 ADDITION, ENDOSKELETAL SYSTEM, BELOW KNEE, ALIGNABLE SYSTEM
- L5920 ADDITION, ENDOSKELETAL SYSTEM, ABOVE KNEE OR HIP DISARTICULATION, ALIGNABLE SYSTEM
- L5925 ADDITION, ENDOSKELETAL SYSTEM, ABOVE KNEE, KNEE DISARTICULATION OR HIP DISARTICULATION, MANUAL LOCK
- L5930 ADDITION, ENDOSKELETAL SYSTEM, HIGH ACTIVITY KNEE CONTROL FRAME
- L5940 ADDITION, ENDOSKELETAL SYSTEM, BELOW KNEE, ULTRA-LIGHT MATERIAL (TITANIUM, CARBON FIBER OR EQUAL)
- L5950 ADDITION, ENDOSKELETAL SYSTEM, ABOVE KNEE, ULTRA-LIGHT MATERIAL (TITANIUM, CARBON FIBER OR EQUAL)
- L5960 ADDITION, ENDOSKELETAL SYSTEM, HIP DISARTICULATION, ULTRA-LIGHT MATERIAL (TITANIUM, CARBON FIBER OR EQUAL)
- L5961 ADDITION, ENDOSKELETAL SYSTEM, POLYCENTRIC HIP JOINT HIP JOINT, PNEUMATIC OR HYDRAULIC CONTROL, ROTATION CONTROL, WITH OR WITHOUT FLEXION AND/OR EXTENSION CONTROL
- L5962 ADDITION, ENDOSKELETAL SYSTEM, BELOW KNEE, FLEXIBLE PROTECTIVE OUTER SURFACE COVERING SYSTEM
- L5964 ADDITION, ENDOSKELETAL SYSTEM, ABOVE KNEE, FLEXIBLE PROTECTIVE OUTER SURFACE COVERING SYSTEM
- L5966 ADDITION, ENDOSKELETAL SYSTEM, HIP DISARTICULATION, FLEXIBLE PROTECTIVE OUTER SURFACE COVERING SYSTEM
- L5968 ADDITION TO LOWER LIMB PROSTHESIS, MULTIAXIAL ANKLE WITH

SWING PHASE ACTIVE DORSIFLEXION FEATURE

L5970 ALL LOWER EXTREMITY PROSTHESES, FOOT, EXTERNAL KEEL, SACH FOOT

L5971 ALL LOWER EXTREMITY PROSTHESIS, SOLID ANKLE CUSHION HEEL (SACH) FOOT, REPLACEMENT ONLY

L5972 ALL LOWER EXTREMITY PROSTHESES, FOOT, FLEXIBLE KEEL

L5973 ENDOSKELETAL ANKLE FOOT SYSTEM, MICROPROCESSOR CONTROLLED FEATURE, DORSIFLEXION AND/OR PLANTAR FLEXION CONTROL, INCLUDES POWER SOURCE.

L4466 GARMENT, BELT, SLEEVE OR OTHER COVERING, ELASTIC OR SIMILAR STRETCHABLE MATERIAL, ANY TYPE, EACH (non-covered)

L5974 ALL LOWER EXTREMITY PROSTHESES, FOOT, SINGLE AXIS ANKLE/FOOT

L5975 ALL LOWER EXTREMITY PROSTHESIS, COMBINATION SINGLE AXIS ANKLE AND FLEXIBLE KEEL FOOT

L5976 ALL LOWER EXTREMITY PROSTHESES, ENERGY STORING FOOT (SEATTLE CARBON COPY II OR EQUAL)

L5978 ALL LOWER EXTREMITY PROSTHESES, FOOT, MULTIAXIAL ANKLE/FOOT

L5979 ALL LOWER EXTREMITY PROSTHESIS, MULTI-AXIAL ANKLE, DYNAMIC RESPONSE FOOT, ONE PIECE SYSTEM

L5980 ALL LOWER EXTREMITY PROSTHESES, FLEX FOOT SYSTEM

L5981 ALL LOWER EXTREMITY PROSTHESES, FLEX-WALK SYSTEM OR EQUAL

L5982 ALL EXOSKELETAL LOWER EXTREMITY PROSTHESES, AXIAL ROTATION UNIT

L5984 ALL ENDOSKELETAL LOWER EXTREMITY PROSTHESIS, AXIAL ROTATION UNIT, WITH OR WITHOUT ADJUSTABILITY

L5985 ALL ENDOSKELETAL LOWER EXTREMITY PROSTHESES, DYNAMIC PROSTHETIC PYLON

L5986 ALL LOWER EXTREMITY PROSTHESES, MULTI-AXIAL ROTATION UNIT ('MCP' OR EQUAL)

L5987 ALL LOWER EXTREMITY PROSTHESIS, SHANK FOOT SYSTEM WITH VERTICAL LOADING PYLON

L5988 ADDITION TO LOWER LIMB PROSTHESIS, VERTICAL SHOCK
REDUCING PYLON FEATURE

L5990 ADDITION TO LOWER EXTREMITY PROSTHESIS, USER
ADJUSTABLE HEEL HEIGHT

L5999 LOWER EXTREMITY PROSTHESIS, NOT OTHERWISE SPECIFIED

L7367 LITHIUM ION BATTERY, REPLACEMENT

L7368 LITHIUM ION BATTERY CHARGER

L7510 REPAIR OF PROSTHETIC DEVICE, REPAIR OR REPLACE MINOR
PARTS

L7520 REPAIR PROSTHETIC DEVICE, LABOR COMPONENT, PER 15
MINUTES

L7600 PROSTHETIC DONNING SLEEVE, ANY MATERIAL, EACH

L7700 GASKET OR SEAL, FOR USE WITH PROSTHETIC SOCKET INSERT,
ANY TYPE, EACH

L8400 PROSTHETIC SHEATH, BELOW KNEE, EACH

L8410 PROSTHETIC SHEATH, ABOVE KNEE, EACH

L8417 PROSTHETIC SHEATH/SOCK, INCLUDING A GEL CUSHION LAYER,
BELOW KNEE OR ABOVE KNEE, EACH

L8420 PROSTHETIC SOCK, MULTIPLE PLY, BELOW KNEE, EACH

L8430 PROSTHETIC SOCK, MULTIPLE PLY, ABOVE KNEE, EACH

L8440 PROSTHETIC SHRINKER, BELOW KNEE, EACH

L8460 PROSTHETIC SHRINKER, ABOVE KNEE, EACH

L8470 PROSTHETIC SOCK, SINGLE PLY, FITTING, BELOW KNEE, EACH

L8480 PROSTHETIC SOCK, SINGLE PLY, FITTING, ABOVE KNEE, EACH

▼ **Important Note:**

Northwood's Medical Policies are developed to assist Northwood in administering plan benefits and determining whether a particular DMEPOS product or service is reasonable and necessary. Equipment that is used primarily and customarily for a non-medical purpose is not considered durable medical equipment.

Coverage determinations are made on a case-by-case basis and are subject to all of the terms, conditions, limitations, and exclusions of the member's contract including medical necessity requirements.

The conclusion that a DMEPOS product or service is reasonable and necessary does not constitute coverage. The member's contract defines which DMEPOS product or service is covered, excluded or limited. The policies provide for clearly written, reasonable and current criteria that have been approved by Northwood's Medical Director.

The clinical criteria and medical policies provide guidelines for determining the medical necessity for specific DMEPOS products or services. In all cases, final benefit determinations are based on the applicable contract language. To the extent there are any conflicts between medical policy guidelines and applicable contract language, the contract language prevails. Medical policy is not intended to override the policy that defines the member's benefits, nor is it intended to dictate to providers how to direct care. Northwood Medical policies shall not be interpreted to limit the benefits afforded to Medicare or Medicaid members by law and regulation and Northwood will use the applicable state requirements to determine required quantity limit guidelines.

Northwood's policies do not constitute medical advice. Northwood does not provide or recommend treatment to members. Members should consult with their treating physician in connection with diagnosis and treatment decisions.

▼ References

Centers for Medicare and Medicaid Services, Medicare Coverage Database, National Coverage Documents; October 2015.

CGS Administrators, LLC. Jurisdiction B DME MAC, Lower Limb Prostheses. Local Coverage Determination No. L33787; revised date October 1, 2015. Accessed December 2017. Reviewed December 17, 2018.

Noridian Healthcare Solutions, LLC. Jurisdiction A DME MAC, Lower Limb Prostheses. Local Coverage Determination No. L33787; revised date October 1, 2015. Accessed December 2017. Reviewed December 17, 2018.

Applicable URAC Standard

Core 8	Staff operational tools and support
--------	-------------------------------------

Change/Authorization History

Revision Number	Date	Description of Change	Prepared /Reviewed by	Approved by	Review Date:
A	Nov.2006	Initial Release	Rosanne Brugnoni	Ken Fasse	n/a
01	Jan.2007	HCPCS Codes: Added L5993, L5994. Revised L5848, L5995	Susan Glomb	Ken Fasse	
02	Oct.2008	Moved: Noncoverage statement for user adjustable heel heights.	Susan Glomb	Ken Fasse	
03		Annual Review / no changes	Susan Glomb	Ken Fasse	Dec.2008
04	Dec.2008	Delete Codes L5993,L5994 and L5995	Susan Glomb	Ken Fasse	
05	12-22-09	Annual Review/ no changes	Susan Glomb	Ken Fasse	Dec.2009
06	01-05-10	Added Code: L5973, endoskeletal ankle foot system, microprocessor controlled feature, dorsiflexion and/or plantar flexion control, includes power source. Added Code A4466 Garment, Belt, sleeve or other covering, elastic or similar stretchable material, any type, ea. (non-covered). E0700 – new narrative- Safety equipment, device or accessory, any type. E0249 – Pad for water circulating heat unit, for replacement only. Discontinued codes: E1340, L0210, L3651, L3652, L3700, L3701, L3909, L3911, L6639.	Susan Glomb	Ken Fasse	
07	12-02-10	Annual Review – No changes	Susan Glomb	Ken Fasse	Dec.2010
08	07-20-11	Added Important Note to all Medical Policies	Susan Glomb	Dr. B. Almasri	
09	12-6-11	Annual Review. Added References to Policy	Susan Glomb	Dr. B. Almasri	Dec. 2011
10	10-22-12	Quantity changes made to L8400, L8410, L8417, L8420, L8430, L8440, L8460 and L8470	Susan Glomb	Dr. B. Almasri	
11	11-29-12	Deleted code: L5311. Added code: L5312.	Susan Glomb	Dr. Almasri	
12	12-03-12	Annual review – no further changes.	Susan Glomb	Dr. B. Almasri	Dec. 2012

13	12-30-13	Annual review. Added L5859 Addition to lower extremity prosthesis, endoskeletal knee-shin system, powered and programmable flexion/extension assist control, includes any type motor(s). Changed narrative to L5972. All lower extremity prostheses, foot, flexible keel.	Susan Glomb	Dr. B. Almasri	
14	12-8-14	Annual Review. Added L5973. Microprocessor controlled ankle foot system (L5973) energy storing foot (L5976), dynamic response foot with multi-axial ankle (L5979), flex foot system, flex-walk system or equal (L5981), or shank foot system with vertical loading pylon (L5987) is covered for members whose functional level is 3 or above.	Susan Glomb	Dr. B. Almasri	
15	12-14-15	Annual Review. References updated.	Susan Glomb	Dr. B. Almasri	12-14-15
16	12-08-16	Annual Review. No Changes.	Lisa Wojno	Dr. B. Almasri	12-08-16
17	12-15-17	Annual Review. Updated DME MAC reference names.	Lisa Wojno	Dr. Cheryl Lerchin	12-15-17
18	12-17-18	Annual review. Added L5859 coverage; coding guidelines for L7700, L5704-L5707 and L5962, L5964, L5966 per Medicare. Added bundling table for batteries/chargers.	Carol Dimech	Dr. C. Lerchin	12-17-18